

Pre-Operative

- Laser door signs/goggles all doors
- Windows covered (holmium, CO2 not required)
- Water in basin
- Foot peddle accessible
- Non-reflective laser instruments
- Laser test fired
- Fire extinguisher in room
- Keys obtained (locked cupboard)
- Appropriate laser goggles (wave length and ocular density checked)
- Metal garbage can with lid
- Smoke evacuator tested
 - Colposcopy – stackhouse smoke evac
 - Laparoscopy – AMT endoscopy smoke evac
 - Holmium Laser – not required

CO2 LASER = TEST FIRE 10 WATTS + 100MJ ULTRAPULSE FOR ALL PROCEDURES

Intra-Operative

- All staff wearing appropriate laser goggles
 - Patient: Asleep – Saline soaked eye pads and goggles
Awake – Appropriate laser goggles
 - All staff wearing laser mask or N95 mask for colposcopy
 - Site draped: wet towels for holmium and colposcopy
 - Water in basin all procedures
 - Non-reflective instruments
-

Post-Operative

- Laser documentation and log completed
- Keys returned
- Laser and accessories cleaned and stored properly
- Laser malfunctions noted and reported
- Laser safety goggles cleaned